

Statement of Conformity

Declaration: ATEX Zone 2 apparatus

Document No. **SC110324E** Issue No.: **2** Page 1 of 1

The electrical apparatus
manufactured by
complies with
and may be marked with

IsoAmp® PWR A 20100 F0
Knick Elektronische Messgeräte GmbH & Co. KG
EN 60079-0: 2012 / A11: 2013
EN 60079-15: 2010
Ex II 3 G Ex nA IIC T6 Gc X

Description	Repeater power supply used for the supply of a 2-wire transmitter via a 4 ... 20 mA current loop with 0/4 ... 20 mA and 0 ... 10 V output.
Electrical Data	<p>Current Loop (terminals 1+, 2-) 4 ... 20 (22) mA, 16.5 V Signal Output (terminals 5-, 6+) 22 mA, 11 V Power Supply (terminals 3-, 4+) 24 VDC (+/-15%), 0.6 W (SELV)</p> <p>The current loop circuit is galvanically separated from the signal output circuit and from power supply up to a peak voltage of 300 V. The signal output circuit and power supply are galvanically separated from each other up to a peak voltage of 60 V.</p> <p>The ambient temperature range is 0 °C ≤ Ta ≤ +55 °C.</p>
Specific Conditions of Use	The IsoAmp® PWR A 20100 F0 must be installed in a housing (control box or distributor box) that fulfils the requirements according to EN 60079-15 as well as at least those according to the IP54 (EN 60529) degree of protection.
Evaluation	<p>The IsoAmp® PWR A 20100 F0 complies with the standards EN 60079-0: 2012 / A11: 2013 and EN 60079-15: 2010. The apparatus marking shall include the code</p> <p>Ex II 3 G Ex nA IIC T6 Gc X</p> <p>The IsoAmp® PWR A 20100 F0 may be used in hazardous areas Zone 2.</p>
Test documents	UL File QVAJ.E340288, Process Control Equipment for Use in Zone Classified Hazardous Locations, Class I, Zone 2.
Place and date of issue	Berlin, May 12th, 2016

Knick Elektronische Messgeräte GmbH & Co. KG


ppa.
Jörg Giebson

Leiter Bereich Forschung und Entwicklung
Vice President R & D