

Certificate of Compliance

Certificate: 1555681

Master Contract: 188909

Project: 1555681

Date Issued: August 18, 2004

Issued to: Knick Elektronische Meßgeräte GmbH & Co.
Postfach 37 04 15
14134 Berlin
GERMANY

*The products listed below are eligible to bear the CSA Mark shown
with adjacent indicators 'C' and 'US'*

A handwritten signature in black ink, appearing to be 'O. Ewanchyna', is written over a horizontal line.

Authorized by: O. Ewanchyna, P. Eng.

PRODUCTS

Class 2252 03 Process Control Equipment
Class 2252 83 Process Control Equipment – Certified to U.S. Standards

Modular measuring system for liquid analysis.
Model: Protos 3400 S (Stainless steel enclosure), Protos 3400 C (Coated steel enclosure)
Input: 24 to 230 V AC/DC, 45 to 65 Hz, 15 VA / 10 W

APPLICABLE REQUIREMENTS

- | | | |
|------------------------|---|---|
| CAN/C22.2 No.1010.1-92 | - | Safety requirements for electrical equipment for measurement, control, and laboratory use, Part 1: General Requirements.
Amendment 2: 1997 to CAN/CSA-C22.2 No 1010.1-92 |
| UL61010C-1 | - | Safety requirements for process control equipment |

CONDITIONS OF ACCEPTABILITY

The above module have been evaluated for use in a max. ambient temperature of 55°C, pollution degree 2 and Installation Category II environment

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Supplement to Certificate of Compliance

Certificate: 1555681

Master Contract: 188909

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1555681	August 18, 2004	Original cCSAus Certification of Modular measuring system for liquid analysis. Model: Protos 3400 S (Stainless steel enclosure), Protos 3400 C (Coated steel enclosure).